

TRANSDUCER VERSATORR

BIBVT100

Brooks Transducer BVT100

- Type: Pirani Transducer
- Temperatur: Ambient -50°C
- Tryk område: 7.5×10^{-7} Torr igennem 1000 Torr
- Digital kommunikation: RS-232 /// RS-485



PRODUKTBESKRIVELSE

Software findes her. <https://www.brooksinstrument.com/en/products/accessories-software/product-software/versatorr-gui>

SPECIFIKATIONER

Funktion	Vakuum måler
Trykområde max	10 bar
Analog output	0.5 - 9.5 Vdc, 1.0-9 Vdc, 0.375 - 5.659 Vdc, 1.9 - 10 Vdc, 1.5 - 8.5 Vdc. 1.9 - 9.1 Vdc, 0-10Vdc, 0-5Vdc
Digital output	RS232, RS485
Temperaturområde fra	-20 °C
Temperaturområde til	50 °C
Tilslutning	Flange DN16KF, DN25KF, VCR8F
Medietemperatur fra	-20 °C
Nøjagtighed	±0.25% of reading
Medietemperatur til	85 °C
Certifikater	UL, CSA, EN/IEC, CE, RoHS
Elektrisk tilslutning	9-Pin SUB-D Male, 15-Pin HD SUB-D Male, 6-Pin Hirschmann, 8-Pin RJ45 / FC68